


TSN A1:D100

Filter: No filter

Last updated 18 oktober 2022 at 17:33 by S.A.A. van den Berg

Descriptives



N | 69

	Minimum	Maximum
LCMS nmo/L	0,1770	47,7840
Atellica nmol/L	0,3000	42,8000

Fit Y on X

Passing-Bablok fit

Equation | Atellica nmol/L = -0,1401 + 1,096 LCMS nmo/L

Parameter	Estimate	Bootstrap 95% CI
Intercept	-0,1401	-0,2333 to 0,04983
Slope	1,096	1,006 to 1,147

CI based on 999 bootstrap samples.